

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

~~JUL 22 2009~~

JUL 16 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130112
Name		Bruce Smith Jr	ID# 3522 Date 7-8-09
A	Agency	Wrangell Police Department	Phone # (907) 874-3304
Instrument Location Wrangell Police Department, 431 Zimovia Highway			
Alco S/N		X172933	Target Value .080 High Pressure 1050
B	Alco Test Values	1 <sup>st</sup> Alco .085	2 <sup>nd</sup> Alco .085
Signature		B. Smith	
(OVER)			

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 24<sup>th</sup> day of July, 2009.

Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130112 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130112

JULY 08, 2009 ✓ *CUS 7/8/09*

OPERATOR'S NAME:

SMITH JR, BRUCE W

OPERATOR'S NUMBER: 3522

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

CAL

O.L. #: 012345

DEPT/AGENCY: WR61

CASE/REPORT: 89070000

TEST TYPE: V

ALCO TARGET VALUE: .000 ✓

ALCO S/N: X172933 ✓

— BREATH ANALYSIS —

.000 ADJUSTED FOR 30.24 in

ALCO TARGET .000 19:50

BLANK TEST .000 19:51

INTERNAL STANDARD VERIFIED 19:51

ALCO TU 30.24 in .005 ✓ 19:51

BLANK TEST .000 19:52

SUBJECT SAMPLE *CUS 7/8/09* .000 19:52

BLANK TEST .000 19:53

ALCO TU 30.25 in .005 ✓ 19:53

BLANK TEST .000 19:54

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130112

JULY 08, 2009 ✓

TIME 19:45

— DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 40c

BREATH TUBE: 43c

BAROMETER: 30.25 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!""#%&'()\*+,-./0123456789:;<=>?@ABCDEF6  
HIJ KLMNOPQRSTU VWXYZ[\]^\_`abode fghijklmno  
pqrstuwxz{|}~